



Electronic copy is controlled under document control procedure. Hard copy is uncontrolled & under responsibility of beholder

It is allowed ONLY to access and keep this document with who issued, who is responsible and to whom it is applicable.

Information Security Classification: Open Confidential Sensitive Secret

Colon Hydrotherapy Inspection Checklist- Final

Name of the Facility: _____

Date of Inspection: ____/____/____

Ref.	Description	Yes	No	N/A	Remarks
2	General Requirements				
2.7.	Specialized plumbing shall be provided for the human waste released during the treatment so that it is flushed directly into the sewage water system of the health facility.				
2.8.	There shall be a sediment filter in addition to an activated carbon filter attached to the hot and cold water connection of the treatment room.				
2.9.	There shall also be a second filter system attached to the colon hydrotherapy device itself in accordance with the manufacturer specification.				
2.10.	The filters shall be changed in a timely manner as recommended by the manufacturer and a filter change record/ log book shall be maintained.				
2.11.	In some cases, specifically the closed pressure systems, the installation of an ultraviolet sterilizer at the main supply shall be provided for water treatment.				
3	Facility Requirements				
3.1.2.	The procedure room or treatment room) shall have:				
3.1.2.1.	A floor that is entirely washable.				
3.1.2.2.	Carpets on the floor shall not be permitted.				
3.1.2.3.	Its own toilet and washing facilities.				
3.1.2.4.	An extraction system.				
3.1.2.5.	A hand-washing station with a hands free operating tap and liquid or foam soap dispensers shall be provided in all				

Checklist	ID	Issue#	Issue Date	Revision Date	Page#
Colon Hydrotherapy / Final	CP_9.6.01_F07	1	Nov 14, 2022	Nov 14, 2025	1/3



	examination room(s). Sinks shall be designed with deep basins, made of porcelain, stainless steel, or solid surface materials.				
3.1.2.6.	Hand sanitizer dispenser shall be provided in addition to hand-washing stations.				
3.1.2.7.	Provisions for hand drying shall be available at all hand-washing stations.				
3.1.2.8.	A shower for exclusive use of the client located in the procedure room or as an ensuite.				
3.1.2.9.	A paper towel on the client couch.				
3.1.2.10.	A paper towel for each client to clean themselves after the irrigation procedure.				
3.1.3.	The outpatient facility shall provide clean and comfortable changing area for the patient.				
3.1.4.	Clean gowns, robes and towels shall be provided.				
3.1.6.	There shall be provision of drinking water.				
3.1.7.	Emergency medication and crash cart trolley shall be available in the procedure room and in accordance with the crash cart specification stipulated in the DHA Outpatient Facility Regulation available on the DHA website www.dha.gov.ae .				
6	Procedure Requirements				
6.1.4.	Patients' health records shall be maintained in a secured area. In case the health facility utilizes electronic patient records, the health facility must ensure the records are safely maintained and easily retrievable.				
6.1.9.	Under no circumstances shall the colonic irrigation equipment be connected directly to a drinking water supply system. This could result in serious, possibly fatal injury to the patient due to application of mains pressure.				
6.2.	Water pressure:				
6.2.1.	Manufacturer's instructions shall be followed.				
6.3.1.	Water temperature shall be regulated to normal body temperature to prevent thermal shock or scalding. The minimum allowed temperature shall be 31°C and the maximum				

Checklist	ID	Issue#	Issue Date	Revision Date	Page#
Colon Hydrotherapy / Final	CP_9.6.01_F07	1	Nov 14, 2022	Nov 14, 2025	2/3



	allowed temperature is 39°C with the average treatments temperature of 34°C to 37°C. It is unsafe to operate the equipment outside the recommended temperature levels.				
7	Infection control				
7.1.	There shall be disinfectant in the room e.g. sterilizing alcohol.				

Checklist	ID	Issue#	Issue Date	Revision Date	Page#
Colon Hydrotherapy / Final	CP_9.6.01_F07	1	Nov 14, 2022	Nov 14, 2025	3/3