



Electronic copy is controlled under document control procedure. Hard copy is uncontrolled & under responsibility of beholder

It is allowed ONLY to access and keep this document with who issued, who is responsible and to whom it is applicable.

Information Security Classification:  Open  Confidential  Sensitive  Secret

## Platelet Rich Plasma (PRP) Inspection Checklist- Final

Name of the Facility: \_\_\_\_\_

Date of Inspection: \_\_\_\_/\_\_\_\_/\_\_\_\_

Ref.	Description	Yes	No	N/A	Remarks
<b>5</b>	<b>Procedure Requirements</b>				
5.3.	PRP should be obtained using a separating device designed for autologous blood.				
5.6.	Informed Consent form must be made available to the patient before the procedure . The consent form must contain all the fields identified in the sample Consent Form ( please refer to <i>Appendix 1</i> ).				
5.8.	The PRP procedure including; blood drawing, centrifugation and administering the final product must be done in the same room where the patient is present.				
<b>7</b>	<b>General Considerations</b>				
7.2.2.	Contract with a specialized and quality approved company to regularly collect, transport and destroy medical waste materials according to the conditions issued by Public Health Department in Dubai Municipality.				

Checklist	ID	Issue#	Issue Date	Revision Date	Page#
Platelet Rich Plasma (PRP) / Final	CP_9.6.01_F26	1	Nov 14, 2022	Nov 14, 2025	1/1